Calibration Test Report

Approver	Eric Bai
Inspector	Eric Bai
Inspector	Eric Bai

Step 1：

Pour part of salt and part of water (purified water or distilled water is required)

Step 2：

Stir evenly to a semi-melted state, and you can see salt particles.(This semi-melted state is the correct saturated salt solution)

Step 3：

Put the saturated salt solution into the container and carry out the sealing calibration measurement for 10-30 minutes.

Most important:
It does not matter if the instrument is turned off during the sealing test. If it is turned off, please do not separate the instrument and the container. The instrument and the container must be sealed together for 10-30 minutes . After 10-30 minutes, the calibrated value will show the correct value when you turn on the instrument .

Calibration reading result:0.76AW

Balance relative humidity

Saturated salt solution

Relative humidity (%RH)			
Temperature (°C)	Lithium chloride	Potassium acetate	Magnesium chloride
0	11.23 ± 0.54		33.66 ± 0.33
5	11.26 ± 0.47		33.60 ± 0.28
10	11.29 ± 0.41	23.28 ± 0.53	33.47 ± 0.24
15	11.30 ± 0.35	23.40 ± 0.32	33.30 ± 0.21
20	11.31 ± 0.31	23.11 ± 0.25	33.07 ± 0.18
25	11.30 ± 0.27	22.51 ± 0.32	32.78 ± 0.16
30	11.28 ± 0.24	21.61 ± 0.53	32.44 ± 0.14
35	11.25 ± 0.22		32.05 ± 0.13
40	11.21 ± 0.21		31.60 ± 0.13
45	11.16 ± 0.21		31.10 ± 0.13
50	11.10 ± 0.22		30.54 ± 0.13
55	11.03 ± 0.23		29.93 ± 0.16
60	10.95 ± 0.26		29.26 ± 0.18
65	10.86 ± 0.29		28.54 ± 0.21
70	10.75 ± 0.33		27.77 ± 0.25
75	10.64 ± 0.38		26.94 ± 0.29
80	10.51 ± 0.44		26.05 ± 0.34
85	10.38 ± 0.51		25.11 ± 0.39
90	10.23 ± 0.59		24.12 ± 0.46
95	10.07 ± 0.67		23.07 ± 0.52
100	9.90 ± 0.77		21.97 ± 0.60

Saturated salt solution
A convenient way to calibrate the humidity sensor is to use a saturated salt solution. At any temperature, the concentration of the saturated solution is fixed and no measurement is required. Providing excess solute, the solution will remain saturated even in the presence of an appropriate amount of water and water vapor. When the solute is in a solid state in the pure phase, it is easy to determine that the solutibn is saturated.
A saturated salt solution is a slurry mixture of distilled water and chemically pure salt, enclosed in a sealed metal or glass container. Wexler
And Hasegawa measured atmospheric humidity with a dew point hygrometer using more than 8 saturated salt solutions (for atmospheric temperatures of 0 to 50 °C). Rear
Greenspan compiled data on 28 saturated salt solutions based on literature, covering the entire range of relative humidity. Using a database obtained from 21 independent surveys (containing 1106 measurements), the “best” atmospheric relative humidity value was obtained as a function of temperature by applying the least squares method to the conventional polynomial, and the measurement results were obtained. These values ??are summarized in the table below.

Relative humidity (%RH)						
Temperature(°C)	Potassium carbonate	Magnesium nitrate	Sodium chloride	Potassium chloride	Potassium nitrate	Potassium sulfate
0	43.13 ± 0.66	60.35 ± 0.55	75.51 ± 0.34	88.61 ± 0.53	96.33 ± 2.9	98.77 ± 1.1
5	43.13 ± 0.50	58.86 ± 0.43	75.65 ± 0.27	87.67 ± 0.45	96.27 ± 2.1	98.48 ± 0.91
10	43.14 ± 0.39	57.36 ± 0.33	75.67 ± 0.22	86.77 ± 0.39	95.96 ± 1.4	98.18 ± 0.76
15	43.15 ± 0.33	55.87 ± 0.27	75.61 ± 0.18	85.92 ± 0.33	95.41 ± 0.96	97.89 ± 0.63
20	43.16 ± 0.33	54.38 ± 0.23	75.47 ± 0.14	85.11 ± 0.29	94.62 ± 0.66	97.59 ± 0.53
25	43.16 ± 0.39	52.89 ± 0.22	75.29 ± 0.12	84.34 ± 0.26	93.58 ± 0.55	97.30 ± 0.45
30	43.17 ± 0.50	51.40 ± 0.24	75.09 ± 0.11	83.62 ± 0.25	92.31 ± 0.60	97.00 ± 0.40
35		49.91 ± 0.29	74.87 ± 0.12	82.95 ± 0.25	90.79 ± 0.83	96.71 ± 0.38
40		48.42 ± 0.37	74.68 ± 0.13	82.32 ± 0.25	89.03 ± 1.2	96.41 ± 0.38
45		46.93 ± 0.47	74.52 ± 0.16	81.74 ± 0.28	87.03 ± 1.8	96.12 ± 0.40
50		45.44 ± 0.60	74.43 ± 0.19	81.20 ± 0.31	84.78 ± 2.5	95.82 ± 0.45
55			74.41 ± 0.24	80.70 ± 0.35		
60			74.50 ± 0.30	80.25 ± 0.41		
65			74.71 ± 0.37	79.85 ± 0.48		
70			75.06 ± 0.45	79.49 ± 0.57		
75			75.58 ± 0.55	79.17 ± 0.66		
80			76.29 ± 0.65	78.90 ± 0.77		
85				78.68 ± 0.89		
90				78.50 ± 1.0		
95						
100						

Junee (Miss)

HeFei Vetus Electronic Technology Co.,Ltd.